

Certificate of Calibration

Customer: CUSTOMER NAME
1234 EAST DR.
NAPERVILLE, IL 60563
888-000-0000

P.O. Number: POA1B2

ID Number: LL9999



Description: PRESSURE TRANSDUCER
Manufacturer: OMEGA
Model Number: PX409-030GI
Serial Number: GP90631
Technician: HANS BEDERKA

Calibration Date: 02/02/2026
Calibration Due: 02/02/2027
Procedure: TMI-PRESSURE GAUGES
Rev: 2/16/2024
Temperature: 68 °F
Humidity: 15 % RH
As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

On-Site Calibration:
Comments:

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/ NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration is within the current Scope of Accreditation and complies with the requirements of ISO/IEC 17025:2017 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.

B. SCHICKOWSKI, BRANCH MANAGER

Scott Chamberlain, QUALITY MANAGER

Calibration Standards

Asset Number	Manufacturer	Model Number	Date Calibrated	Cal Due
RFD512	ADDITEL CORPORATION	ADT681-02-GP30-PSI-N	11/11/2025	7/5/2026



Technical
Maintenance
Incorporated
A TruMetric Family

3248 FOREST VIEW RD, ROCKFORD, IL 61109

Phone: 779-774-3877

www.tmicalibration.com



CERT #2348.01

Certificate of Calibration

Data Sheet

<u>Parameter</u>	<u>Nominal</u>	<u>Minimum</u>	<u>Maximum</u>	<u>As Found</u>	<u>As Left</u>	<u>Uncertainty +/-</u>	<u>Unit ADJ/FAIL</u>
Pressure Accuracy	3.000	2.970	3.030	2.998	2.998	0.0078 psi	psia
Pressure Accuracy	9.000	8.970	9.030	9.003	9.003	0.0078 psi	psia
Pressure Accuracy	23.000	22.970	23.030	23.005	23.005	0.0078 psi	psia
Pressure Accuracy	27.000	26.970	27.030	27.024	27.024	0.0078 psi	psia
Pressure Accuracy,Hysteresis	27.000	26.970	27.030	26.998	26.998	0.0078 psi	psia
Pressure Accuracy,Hysteresis	23.000	22.970	23.030	23.009	23.009	0.0078 psi	psia
Pressure Accuracy,Hysteresis	9.000	8.970	9.030	9.012	9.012	0.0078 psi	psia
Pressure Accuracy,Hysteresis	3.000	2.970	3.030	3.015	3.015	0.0078 psi	psia

SAMPLE



**Technical
Maintenance
Incorporated**
A TruMetric. Family

3248 FOREST VIEW RD, ROCKFORD, IL 61109

Phone: 779-774-3877

www.tmiclibration.com

